



EC DECLARATION OF CONFORMITY

Name and address of the manufacturer: **CHANGZHOU XIANGRONG MEDICAL INSTRUMENTS CO.,D.**
5 XINMA RD,LUOYANG
TOWN,CHANGZHOU,JIANGSU,PROVICE,CHINA

Name and address of the European Authorized Representative **ICR POLSKA CO., LTD**
UL.PLAC PRZYMIERZA 6.03-944 WARSZAWA
WWW.ICRPOLSKA.COM

We declare under our sole responsibility that

the device: **Disposable Masks**

according to EN 149:2001+A1:2009

meets the provisions of the CE Directive(s): R 2016/425 (Personal Protective Equipment)
and its transpositions in national laws which apply to it.

Conformity assessment procedure: **EN 149:2001+A1:2009**

Registration No.: **LW-011702-PPE**

Notified Body: **ICR POLSKA CO., LTD**
UL.PLAC PRZYMIERZA 6.03-944 WARSZAWA
WWW.ICRPOLSKA.COM

Changzhou, April 10th, 2020

Position: International Dept. Manager





北自所(常州)科技发展有限公司

产品质量分析证明书

RIAMB (CHANGZHOU) Science and Technology Co., Ltd.

Certificate of Analysis

RIAMB (CHANGZHOU) Science and Technology Co., Ltd.											
Certificate of Analysis											
客户名称 Customer Name	产品描述 Specification		规格/型号 Type /Specification			订单号 Purchase Order	产品代码 Code Number		生产日期 Produce Date		
向荣 Xiangrong	熔喷无纺布 (过滤材料) Melt-blown non-woven fabric (filter material)		25g/m² × 175mm × 2600m			/	/		2020.04.13		
产品性能 Property	克重 Basis Weight (GSM)	纵向强力 MD-Tensile Strength (N/5cm)	纵向MD伸长率 MD-Elongation (%)	横向强力 CD-Tensile Strength (N/5cm)	横向伸长率 CD-Elongation (%)	过滤效率 Particle filtration efficiency	静水压 Hydrostatic Head (mm H2O)	透气性 Breathability (mm/s)	液体穿透时间 Penetration Time(S)		
						NaCl (2.5um)					
样品尺寸 Sample Size	100cm²	50mm*200mm	50mm*200mm	50mm*200mm	50mm*200mm	100mm*100mm	200mm*200mm	100mm*100mm	125mm*125mm		
样品数量 Sample Number	10	10	10	10	10	5	6	10	1st	2nd	3rd
最小值 Min	24.0	17.4	20.3	11.4	30.6	96.3	275	212	/	/	/
最大值 Max	25.7	20.7	47.0	13.1	55.8	99.7	405	230	/	/	/
平均值 Average	24.7	19.0	33.4	12.5	45.8	97.8	348	222	/	/	/
测试方法 Test Method	ASTM D3776	ASTM D5035	ASTM D5035	ASTM D5035	ASTM D5035	YYT 0969-2013	AATCC 127	ASTMD737	EDANA 150.5-02		
备注 Remark	<div>常州荣兴科技发展有限公司</div>										

编制: 顾文豪

日期: 20.04.14.

审核:

日期:



批准:

日期:

2020.04.14



160010260248



(2016)国认监认字(099)号



中国认可
国际互认
检测
TESTING
CNAS L1499

检 验 报 告

检验报告编号: No. 2020(尘) - 1401

样品名称: 一次性防护口罩

委托单位: 常州市向荣医疗器械有限公司

检验类别: 委 托

国家劳动保护用品质量监督检验中心(北京)




国家劳动保护用品质量监督检验中心（北京）

检 验 报 告



检验报告编号: No. 2020(尘) - 1401

共 2 页 第 1 页

样品名称	一次性防护口罩	商 标	_____
规格型号	_____	防护级别	_____
委托单位名称	常州市向荣医疗器械有限公司	委托单位电话	13861282768
委托单位地址	江苏省常州市武进区新马路 5 号		
生产单位名称	常州市向荣医疗器械有限公司	生产单位电话	13861282768
生产单位地址	江苏省常州市武进区新马路 5 号		
样品数量	10 个	生产日期	2020. 03. 16
样品状态	完好	到样日期	2020. 03. 18
样品特征	蓝色平面口罩	送 样 者	陈海霞
检验依据	GB/T 32610-2016《日常防护型口罩技术规范》		
检验项目	内在质量——吸气阻力、内在质量——呼气阻力、过滤效率		
检 验 结 论	<p>该样品经检验，依据 GB/T 32610-2016《日常防护型口罩技术规范》，判定所检项数据符合标准要求。</p> <div style="text-align: center;">  <p>签发日期: 2020 年 3 月 24 日</p> </div>		
备 注	样品未做预处理		
<p>批 准: 陈海霞 审 核: 冯 强 主 检: 周 芸芸</p>			

检验报告 (检验结果汇总)

共 2 页 第 2 页

A black and white photograph of a rectangular surgical mask with two ear loops, lying flat on a dark surface. The mask has a pleated center section and a non-pleated border. The ear loops are made of a thin, light-colored material. The mask is oriented horizontally.

结 束

检验报告专用章



160010260248



(2016) Guo Ren Jian
Ren Zi No. (099)



China Accredited
International mutually
accredited testing
TESTING
CNAS L1499

Test Report

Test Report No.: No. 2020 (Dust) - 1401

Sample name: Disposable protective mask

Entrusting unit: Changzhou Xiangrong Medical Instruments Co.,Ltd

Test category: Entrusted test



National Quality Supervision and Inspection Center for Labor Protection Supplies (Beijing)

National Quality Supervision and Inspection Center for Labor Protection Supplies (Beijing)

(seal)

National Quality Supervision and Inspection Center for Labor Protection



Supplies (Beijing)

Test Report

Test report No.: No. 2020 (Dust) - 1401

Page 1 of 2

Sample name	Disposable protective mask	Trademark	-----
Model and specification	-----	Protection grade	-----
Name of entrusting unit	Changzhou Xiangrong Medical Instruments Co.,Ltd	Tel. of entrusting unit	13861282768
Address of entrusting unit	No. 5, Xinma Road, Wujin District, Changzhou City, Jiangsu Province		
Name of production unit	Changzhou Xiangrong Medical Instruments Co.,Ltd	Tel. of production unit	13861282768
Address of production unit	No. 5, Xinma Road, Wujin District, Changzhou City, Jiangsu Province		
Number of samples	10	Production date	March 16, 2020
Sample status	Perfect	Sample arrival date	March 18, 2020
Sample feature	Blue plane mask	Sample sender	Chen Haixia
Test basis	GB/T 32610-2016 Technical Specification for Daily Protective Masks		
Test items	Inherent quality - inhalation resistance, inherent quality - exhalation resistance and filtration efficiency		
Test conclusion	<p>The samples underwent the test, according to GB/T 32610-2016 Technical Specification for Daily Protective Masks, the data of the test items is judged to conform to the requirements of the standard.</p> <p>Special Seal for Test Report of National Quality Supervision and Inspection Center for Labor Protection Supplies (Beijing) (seal) (Special seal for test report) Date of issue: March 24, 2020</p>		
Remarks	The samples are not given pre-treatment.		
<p>Approved by: (signature) Reviewed by: (signature) Main tester: Zhou Yunyun(signature)</p>			




National Quality Supervision and Inspection Center for Labor Protection Supplies

(Beijing)

Test Report (Summary of Test Result)

Test report No.: No. 2020 (Dust) - 1401

Page 2 of 2

Serial No.	Test item	Standard clause No.	Standard requirements	Actually tested result				Conclusion of this item	Remarks
1	Inherent quality Inhalation resistance	5.3	≤175 Pa	Without pre-treatment		9#	24 Pa	Conform to the requirements of the standard	-----
						10#	24 Pa		
2	Inherent quality Exhalation resistance	5.3	≤145 Pa	Without pre-treatment		9#	18 Pa	Conform to the requirements of the standard	-----
						10#	18 Pa		
3	Filtration efficiency	5.4	Salty media: Grade I≥99% Grade II≥95% Grade III≥90% Test gas flow: Single filter element (85±4) L/min	Salty media	Without pre-treatment	1#	95.5%	Conform to the requirements of the standard (Grade III)	-----
						2#	96.9%		
						3#	95.5%		
						4#	95.7%		
						5#	95.6%		
			Oily media Grade I≥99% Grade II≥95% Grade III≥80% Test gas flow: Single filter element (85±4) L/min	Oily media	Without pre-treatment	6#	83.3%		
						7#	83.4%		
						8#	83.2%		
						9#	83.2%		
						10#	83.2%		
Environment temperature: (25±5)°C Relative humidity: (30±10)%			Environment temperature: 25°C Relative humidity: 31%						
Photo of sample									
Remarks			-----						

Test date: From March 18, 2020 to March 23, 2020

----- End -----



Business License (Duplicate)

Unified Social Credit Code: 91440300788349389W

Company Name: Shenzhen ODB Translation Co., Ltd.

Type of Business: Limited Liability Company


Place of Business: 15D#, Block A, Zhenye Building, No. 2014 Bao'an Road South, Luohu District, Shenzhen

Legal Representative: Huang Rongfa

Date of Incorporation: May 12, 2006

Important Notes

1. Business scope of the commercial subject is determined by its articles of association. Any item as specified in the business scope and requiring examination and approval as stipulated by laws and regulations cannot be engaged until all required approval documents have been granted.
2. In terms of business scope, licensed and approved business items of the commercial subject as well as its annual financial reports and other credit information, please visit the commercial subject credit information publicity platform provided by the Market and Quality Supervision Commission of Shenzhen Municipality (<http://www.szcredit.com.cn>) or scan QR Code as below for query.
3. The commercial subject should submit its annual report for the previous year to its business registration authorities from January 1st to June 30th of every year. The commercial subject should disclose its information to the society in accordance with the provisions of the *Interim Regulation on Enterprise Information Disclosure*.



Registration Authority:
Seal: Market Supervision Administration Bureau of Shenzhen Municipality

November 08, 2018

Translators Association of China

Certificate of Member

Category of Member: Entity Member

Member Name: Shenzhen ODB Translation Co., Ltd.

Issue Time: January 2019

Uniform Number: DD170819



Translators Association of China
January 2, 2019
(Valid to December 31, 2020)
Translators Association of China (Seal)